

| File No.GEM/TP/Basti Yantra (Anuasana & Mathra)/2019 (Version 3.0) | | | | | | | | Draft T/P | |
|--|---|------|---------------|-----------------------------------|-----|--|------|-------------------|---------|
| Level 1 Category : Medical Equipment and Accessories and Supplies - UNSPS Code 42000000 | | | | | | | | | |
| Level 2 Category : Ayurvedic Equipments and other Related Products - UNSPS Code 42990000 | | | | | | | | | |
| Level 3 Category : Panchakarma Products and Accessories - UNSPS Code 42999900 | | | | | | | | | |
| Level 4 Category: Basti Yantra (Anuasana & Mathra) For Panchakarma - UNSPS Code 42999910 | | | | | | | | GEM Version 3.0 | |
| S.NO | PARAMETER | HINT | TYPE OF FIELD | VALUES | | | UNIT | CATEGORY OF FIELD | REMARKS |
| GENERAL FEATURES | | | | | | | | | |
| 1 | Product Description | -- | ENUMERABLE | Basti Yantra (Anuasana & Mathra) | | | -- | MANDATORY | |
| 2 | Utility | -- | ENUMERABLE | For Panchakarma of all age groups | | | -- | MANDATORY | |
| VASTI YANTRA (ANUVASANA) SPECIFICATIONS | | | | | | | | | |
| 1 | Material | -- | ENUMERABLE | Chrome plated brass body | | | -- | MANDATORY | |
| 2 | Adapter CP socket provided for fixing the screwable and disposable vasti nethrams made of surgical plastic | -- | ENUMERABLE | Yes | | | -- | MANDATORY | |
| 3 | Size of Adapter CP socket | -- | MEASURABLE | 12 | | | mm | MANDATORY | |
| 4 | Hole Size of Tip for Ayurvedic use | -- | MEASURABLE | 5 | | | mm | MANDATORY | |
| 5 | Capacity of Vasti Yantra for Sneha vasti procedures | -- | MEASURABLE | 200 | 250 | | ml | GOLDEN | |
| 6 | Diameter of Vasti Yantra (Anuvasana) | -- | MEASURABLE | 1.5 | | | inch | MANDATORY | |
| 7 | Length of Vasti Yantra (Anuvasana) | -- | MEASURABLE | 9 | | | inch | MANDATORY | |
| VASTI YANTRA (MATRA VASTI) SPECIFICATIONS | | | | | | | | | |
| 1 | Material | -- | ENUMERABLE | Chrome plated brass body | | | -- | MANDATORY | |
| 2 | Adapter CP socket provided for fixing the screwable and disposable vasti nethrams made of surgical plastic | -- | ENUMERABLE | Yes | | | -- | MANDATORY | |
| 3 | Size of Adapter CP socket | -- | MEASURABLE | 12 | | | mm | MANDATORY | |
| 4 | Hole Size of Tip for Ayurvedic use | -- | MEASURABLE | 5 | | | mm | MANDATORY | |
| 5 | Capacity of Vasti Yantra for Mathra Vasti procedure | -- | MEASURABLE | 90 | 100 | | ml | GOLDEN | |
| 6 | Diameter of Vasti Yantra (Matra Vasti) | -- | MEASURABLE | 1.5 | | | inch | MANDATORY | |
| 7 | Length of Vasti Yantra (Matra Vasti) | -- | MEASURABLE | 5 | | | inch | MANDATORY | |
| PACKING | | | | | | | | | |
| 1 | The product should be packed with all its accessories in such a way that there is no transit damage takes place | -- | ENUMERABLE | Yes | | | -- | MANDATORY | |
| 2 | Packing Type (including user instructions) | -- | ENUMERABLE | AS REQUIRED BY BUYER | | | -- | MANDATORY | |

CERTIFICATIONS

| | | | | | | | | | |
|----------|---|----|------------|-----|-------|----|----|-------------------|--|
| 1 | ISO Certified Manufacturer (Proof of the same to be submitted to buyer on demand if ISO Certified) | -- | ENUMERABLE | Yes | No | | -- | MANDATORY, FILTER | |
| 2 | Product Certifications (Proof of the same to be submitted to buyer on demand if certification is available) | -- | ENUMERABLE | BIS | EU-CE | NA | | MANDATORY, FILTER | |

(A. KESAVA RAO)
AD, GEM-SPV

(MANJU SHARMA)
DY.CEO

ACEO