



12	Plate has current limiting nature for safe guarding the patient fom burns when the contact between patient and plate reduces	P	E	Yes										
13	Patient plates shall be usable with Argon beam coagulator	P	E	Yes										
14	Patient Plate can be folded without damaging	P		Yes										
15	Warranty in years without any link to number of surgeries conducted	P	E	1	2	3								
16	Cerifications for product	C	E	USFDA	CE	ICMED	BIS							
17	Certification date and number	C	T											
18	IECE Certification 60601-1:2005 medical electrical equipment Pat 1 General requirements for safey and essentia! performance for safety	C	E	Yes										
18	IEC 60601-2-2:2009 Medical electrical equipment Part 2 ,particular requirements for safey of high frequency surgical equipment	C	E	Yes										



Shahzad

Astt. Director-II(Q/A)

Shri A. V. Muralidharan

DCEO -GeM



tbd : to be declared by vendor

Shahzad  
Astt. Director-II(Q/A)

Shri A. V. Muralidharan  
DIRECTOR -QA

