

6	Material for I.V. Rod	M	E	SS	MS	Al-alloy											
7	Material for bowl	M	E	SS													
8	Material for waste receptable bucket	M	E	SS	Plastic												
9	Material for delivery Bed Mattress	M	E	Latex Foam	PU Foam	Rexin											
Dimensions																	
1	Maximum Adjustable Back Rest Angle in Degree	D	E	30	40	50	60										degree
2	Maximum Trendlenburg Angle in Degree	D	E	10	20	30	40										degree
3	Maximum Reverse Trendlenburg Angle in Degree	D	E	10	20	30	40										degree
4	Maximum Adjustable bed height in mm with 5% tolerance	D	E	300	350	400	450	500	550	600							mm
5	Diameter of Castor Roller in mm	D	E	100	110	125	135	145	150								mm
6	Clearance between Bed Base frame and Floor surface in cm	D	E	10	12	15	18	20	25								cm
7	Maximum thickness of Mattress in mm with tolerance of plus minus 2mm	D	E	65	85	105	125										mm
8	Approx foam Density in kg/m3 tolerance	D	E	30	40	50											kg/m3
10	Therapeutic Weight limit for mattress in kg	D	E	125	150	175	200	225	250								kg
	Length range of delivery bed in mm	D	E	1600-1700	1700-1800												
11	Width range of delivery bed in mm	D	E	500-550	551-600												mm
12	Weight of delivery Bed in Kg	D	N														Kg
Reports																	
1	Availability of ISO certification with manufacturer	R	E	Yes	No												
2	ISO certification No. and Date if ISO certificate not available then put NA	R	T														
3	Copies of test Reports/certificates and other certifications to be furnished to buyer on demand at time of supplies	R	E	Yes	N/A												
Mr.Ajmer Singh (TA)									A V Muralidhran (Dy CEO)								

