

File No.GEM/TP/Dharapathy with Hi-Lo -Tilt Table For Panchakarma /2019 (Version 3.0)							Finalized T/P	
Level 1 Category : Medical Equipment and Accessories and Supplies - UNSPS Code 42000000								
Level 2 Category : Ayurvedic Equipments and other Related Products - UNSPS Code 42990000								
Level 3 Category : Panchakarma Products and Accessories - UNSPS Code 42999900								
Level 4 Category: Dharapathy with Hi-Lo -Tilt Table For Panchakarma - UNSPS Code 42999926							GEM Version 3.0	
S.NO	PARAMETER	HINT	TYPE OF FIELD	VALUES	UNIT	CATEGORY OF FIELD	REMARKS	
<b>GENERAL FEATURES</b>								
1	Product Description	--	ENUMERABLE	Dharapathy With Hi-Lo -Tilt Table For Panchakarma	--	MANDATORY		
3	Purpose	--	ENUMERABLE	To enable comfortable execution of Nasya, Tharpana, Shirodhara, Kaya sekam, Vasti and Pizhichil procedures by tilting of the dharapathi towards head end and leg end as required and enable easy access to the bed top by invalid/ paraplegic/ aged obese patients by the Hi-Lo function	--	MANDATORY		
4	Utility	--	ENUMERABLE	For Panchakarma of all age group	--	MANDATORY		
<b>DHARAPATHY SPECIFICATIONS</b>								
1	Material of Dharapathy	--	ENUMERABLE	NPG -ISO resin and Advance glass fibre	--	MANDATORY		
2	Non-slip profiled surface for the body	--	ENUMERABLE	Yes	--	MANDATORY		
3	Shining smooth surface for the head areas	--	ENUMERABLE	Yes	--	MANDATORY		
4	Dharapathy should be made in traditional shape with an integral catch bowl and drain channel at the head and leg ends	--	ENUMERABLE	Yes	--	MANDATORY		
5	Wood design surface finish having Shirodhara bowl and drain channel	--	ENUMERABLE	Yes	--	MANDATORY		
6	Should be made in sandwich technique with FRP beam stiffeners bonded between 4 mm thick top panel and 2 mm thick bottom panel	--	ENUMERABLE	Yes	--	MANDATORY		
7	PUF reinforcement provided to resist impact during handling, transportation or shipping	--	ENUMERABLE	Yes	--	MANDATORY		
8	Minimum Flexural Modulus of Dharapathy shall be 10000 N/mm	--	ENUMERABLE	Yes	--	MANDATORY		
9	Number of foldable handles provided (Nos)	--	ENUMERABLE	4	--	MANDATORY		
10	Overall Dimensions of Dharapathy (L x W x H) (±5%)	--	ENUMERABLE	287 cm x 78 cm x 14 cm	--	MANDATORY		
11	Bowl Dimensions (dia x depth) (±5%)	--	ENUMERABLE	25 cm x 27.5 cm x 5 cm	--	MANDATORY		
12	Drain Channel Dimensions(L x W) (±5%)	--	ENUMERABLE	30 cm x 10 cm	cm x cm	MANDATORY		
<b>HI-LO TILT TABLE SPECIFICATIONS</b>								
1	Table should consists of a simple yet efficient mechanism to raise and lower the table to near ground level to allow oiled patients, obese persons, paraplegics and invalid patients	--	ENUMERABLE	Yes	--	MANDATORY		

2	Fabricated in rectangular top frame of 220 x 62 cm using 50mm x 25 mm square pipes on supporting base frame made of 50 mm x 25 mm x 16 gauge ERW tubes and 32 cm dia 16 gauge round tubes	--	ENUMERABLE	Yes			--	MANDATORY	
3	Provided with Rubber bushes (32 mm) at the leg support	--	ENUMERABLE	Yes			--	MANDATORY	
4	Provided with two separate cranks for Trendelenburg and hi-lo positions	--	ENUMERABLE	Yes			--	MANDATORY	
5	Provided with tee sections (40 mm x 6mm) for the corner supporters,20-25 micron zinc plated 12.5 mm double nuts and bolt with cap nuts, crank threads and bushes and powder coated finish of 80-110 microns	--	ENUMERABLE	Yes			--	MANDATORY	
6	Provided with a stopper at 75 cm Height to prevent overlift damage	--	ENUMERABLE	Yes			--	MANDATORY	
7	All nuts and bolts should be galvanized	--	ENUMERABLE	Yes			--	MANDATORY	
8	Hand wheels should be connected to 22 mm shafts with square thread	--	ENUMERABLE	Yes			--	MANDATORY	
9	Height adjustment Range	--	ENUMERABLE	45 cm to 75 cm using a hand wheel			--	MANDATORY	
10	Table should be tiltable upto 30 degree using another wheel	--	ENUMERABLE	Yes			--	MANDATORY	
<b>PACKING</b>									
1	The product should be packed with all its accessories in such a way that there is no transit damage takes place	--	ENUMERABLE	Yes			--	MANDATORY	
2	Packing Type (including user instructions)	--	ENUMERABLE	AS REQUIRED BY BUYER			--	MANDATORY	
<b>CERTIFICATIONS</b>									
1	ISO Certified Manufacturer (Proof of the same to be submitted to buyer on demand if ISO Certified)	--	ENUMERABLE	Yes	No		--	MANDATORY, FILTER	
2	Product Certifications (Proof of the same to be submitted to buyer on demand if certification is available)	--	ENUMERABLE	BIS	EU-CE	NA		MANDATORY, FILTER	

(A. KESAVA RAO)  
AD, GEM-SPV

(MANJU SHARMA)  
DY.CEO

ACEO