

26	Display for power	P	E	individual display for Bipolar and Monopolar cut and Monopolar coagulation										
27	RF leakage current in mA	P	N						<150					
28	Availability of trolley with lockable Castor wheel.	A	E	Yes	No									
29	Power supply	p	E	230V (+ / - 15%) 50Hz										
30	Tropicalization: operating Temp and humidity	p	E	upto 40Deg. C and relative humidity up to 85%										
31	Monopolar Foot Switch supplied	A	E	1	2	Not supplied								
32	Bipolar Foot Switch	A	E	1	2	Not supplied								
33	Reusable hand switching pencil	A	E	1	2	Not supplied								
34	Reusable patient plate	A	E	1	2	Not supplied								
35	Bipolar forceps	A	E	1	2	Not supplied								
36	Forceps cord	A	E	1	Not supplied									
37	universal adaptor	A	E	1	2	Not supplied								
38	Leep Electrodes	A	E	1	2	Not supplied								
39	Overall dimensions in mm x mm x mm	D	Text										mm x mm x mm	
40	Weight in kg									<7.5			kg	
41	Instruction Manuals 3 set	P	E	Yes										
42	Warranty	P	E	3	4	5			G					

43	Availability of test report from central govt/ NABL/ILAC accredited lab covering all parameters	R	Ch	Yes	No								
44	Test report number if test report not available put NA	R	Ch										
45	Test report date if test report not available put NA	R	Ch										
46	Name of test lab if test report not available put NA	R	Ch										
47	Address of test lab if test report not available put NA	R	Ch										
48	Copies of all certifications and reports to be provided to buyer on demand at time of supplies	R	Ch	Yes									

Ms. Ritika ,Technical Assistant

Shri A. V. Muralidharan

GeM

DCEO -

Shahzad
Asth. Director-II(Q/A)

Shri A. V. Muralidharan

GeM

DCEO -

tbd : to be declared by vendor

Shahzad
Astt. Director-III(Q/A)

Shri A. V. Muralidharan
DIRECTOR -QA

