

Level 1 Category:: Medical Equipment & Accessories & Supplies UNSPSC:42

Level 2 Category::Emergency & field medical services products UNSPSC:4217

Level 3 Category:: Emergency Medical Services extricating & immobilizing & transporting products : UNSPSC 421716

Level 4 Category: Foldable and Floatable Stretcher (Adult Size) or Air evacuation stretchers or accessories UNSPSC:42171601

S No	Parameters	CH/N	Value 1	Validation	Unit
Generic					
1	Size suitability	CH	For rescue of Adult victim	M	
2	Foldable and Floatable Stretcher Application	CH	Emergency and Rescue Operation work for Natural or Man made Disaster such as Floods, Fire, Earthquake, Riot , Accident etc	M	
3	Dimension when stretcher is open (L x W x H)	CH	78" x 38" x 1.5"	M	inch
4	Dimension when stretcher is close (L x W x H)	CH	24"x 18"x 6"	M	inch
Constructional					
1	Constructional Material Foldable and Floatable Stretcher	CH	Nylon 66, 250 GSM/300 GSM ATY cordura fabric Antifungal, UV stabilised, water resistant, Anti fraying, breathable and washable.	M	
2	Floating material	CH	Heatlon (laminated EPE foam)	M	
3	Handles Counts	N	8 No.	M	Nos
4	Handle Material	CH	38 mm wide Nylon Belt	M	
5	Tensioner Belt	CH	Adjustable Tighten belt with clamp.	M	
6	Length of Nylon Rope with hook attached in cover bag for draw or pull victims	N	12	G	metre
7	Diameter of Nylon Rope	N	≥ 8	M	mm
8	Breaking Strength of Nylon Rope minimum	N	≥1000	M	Kg
9	Covering Bag	CH	Water proof covering bag with chain locking facility for victims belonging and Safety Items	M	
10	Weight of the Foldable and Floatable Stretcher	N	≤3 Kg	M	kg
Performance					
1	Load carrying capacity Maximum	N	150	M	Kg
2	Flotable capacity Maximum	N	150	M	Kg

3	Lifting Capacity Maximum	N	150	M	Kg
4	Retro Reflective Properties	M	Yes ,High visibility Luminus strips are attached for reflection	M	
Certification					
1	Availability of Test Report from Central Govt/ NABL/ILAC accredited lab to prove confirmity of product for declared parameters	CH		M	
2	Name and address of lab/certifying authority	CH		M	
3	Test report No & date	CH		M	
4	Test report to be submitted to the buyer on demand	CH	yes	M	
5	Warranty	N	10	M	years