

FILE No.																
Creator : Mr. Firoj				Moderator : Shri A.V.Muralidharan (Dy. CEO GeM)				Approver : A.K Kamra								
General Purpose and research Centrifuge non refrigerated																
: UNSPSC Code																
S.No	Parameters			Value 1	Value2	Value 3	Value 4	Value 5	Value 6	Value 7	Validation	Golden parameter	Remarks	Priority	Unit	Demand Aggregation
Standards																
1	Confirmity to standards	S	E	European CE (Notified Body)With 4 Digit Notified Body No.	BIS	CE						G				
2	Certification number and date															
3	Confirmity to Quality Management Sytems for Medical Devices	S	E	ISO 13485								G				
4	Certifications Manufacturer			ISO 9001:2015												
5	Manufacturer certification date and number															
6	Confirmity to medical electrical safety standards	S	E	IEC 60601-1 -2, IEC 61010-3-20, IEC 61010-2 -20								G				
General Parameters																
7	Purpose	P	E	To Centrifuge different kinds of samples in different kinds of Testing Labs and research Labs at room Temperature												
8	Centrifuge design	P	E	Bench Top	Floor Model											
9	Assembly Material for manufacture of General Purpose Lab Centrifuge	M	E	Die cast Aluminum, Aluminum sheet, Stainless Steel, CRC Stell,Plastic and Rubber								G				
10	Lid Material of General Purpose Lab Centrifuge	M	E	CRC Steel finish with powder coating or stove painting	ABS Plastic of injection moulding											
11	Body material of General Purpose Lab Centrifuge	M	E	C.R.C Steel Finish with Powder Coating/ Stove Painting.								G				
12	Front Panel of General Purpose Lab centrifuge	M	E	CRC Steel finish with powder coating or stove painting	ABS Plastic of injection moulding							G				

