

File No.GEM/TP/Infrared Lamp For Panchakarma/2019 (Version 3.0)								Finalized T/P		
Level 1 Category : Medical Equipment and Accessories and Supplies - UNSPS Code 42000000										
Level 2 Category : Ayurvedic Equipments and other Related Products - UNSPS Code 42990000										
Level 3 Category : Panchakarma Products and Accessories - UNSPS Code 42999900										
Level 4 Category : Infrared Lamp For Panchakarma - UNSPS Code 42999928								GEM Version 3.0		
S.NO	PARAMETER	HINT	TYPE OF FIELD	VALUES				UNIT	CATEGORY OF FIELD	REMARKS
GENERAL FEATURES										
1	Product Description	--	ENUMERABLE	Infrared Lamp For Panchakarma				--	MANDATORY	
2	Purpose	--	ENUMERABLE	For use in cold, tension, muscle ache etc				--	MANDATORY	
PRODUCT INFORMATION										
1	Product should have adjustable screen with 5 tilt positions	--	ENUMERABLE	Yes				--	MANDATORY	
2	Simple bulb and exclusive design for use in Panchakarma	--	ENUMERABLE	Yes				--	MANDATORY	
3	Power	--	MEASURABLE	100	150			Watt	GOLDEN	
4	Voltage Requirements	--	ENUMERABLE	220-230 V @ 50 Hz				--	MANDATORY	
PACKING										
1	The product should be packed in such a way that there is no transit damage takes place	--	ENUMERABLE	Yes				--	MANDATORY	
2	Packing Type (including user instructions)	--	ENUMERABLE	AS REQUIRED BY BUYER				--	MANDATORY	
CERTIFICATIONS										
1	ISO Certified Manufacturer (Proof of the same to be submitted to buyer on demand if ISO Certified)	--	ENUMERABLE	Yes	No			--	MANDATORY, FILTER	
2	Product Certifications (Proof of the same to be submitted to buyer on demand if certification is available)	--	ENUMERABLE	BIS	EU-CE		NA	--	MANDATORY, FILTER	
3	Safety certifications specific to the device (Proof of the same to be submitted to buyer on demand)	--	ENUMERABLE	IEC 60601-1:2012	IEC 60601-1-2:2014	IEC 60335	IS:13450(Part 1)/2008	--	MANDATORY, FILTER	Multi Select
WARRANTY										
1	Warranty (years)	--	ENUMERABLE	1	2	3		--	MANDATORY, FILTER	

(A. KESAVA RAO)
AD, GEM-SPV

(MANJU SHARMA)
DY.CEO

ACEO