

File No.GEM/TP/ Point of Care Diagnostic Test Kit - Rotavirus Antigen Rapid Test Kit/2018 (Version 3.0)										DRAFT T/P	
Level 1 Category : Laboratory and Measuring and Observing and Testing Equipment - UNSPS Code 4100000											
Level 2 Category : Measuring and observing and testing instruments - UNSPS Code 41110000											
Level 3 Category : Patient point of care testing supplies and equipment - UNSPS Code 41116200											
Level 4 Category : Point of Care Diagnostic Test Kit - Rotavirus Antigen Rapid Test Kit - UNSPS Code 41116295										GEM Version 3.0	
S.NO	PARAMETER	HINT	TYPE OF FIELD	VALUES					UNIT	CATEGORY OF FIELD	REMARKS
GENERAL FEATURES											
1	Product Description	--	ENUMERABLE	Rotavirus Antigen Rapid Test Kit					--	MANDATORY	
2	Clinical Purpose	--	ENUMERABLE	To detect Rotavirus antigen in Stool Samples for diagnosis of Gastroenteritis					--	MANDATORY	
3	Type of Test	--	ENUMERABLE	Qualitative	Quantitative				--	MANDATORY, FILTER	
4	Testing Principle	--	ENUMERABLE	Immuno-chromatographic Principle					--	MANDATORY	
5	Specimen Required for testing	--	ENUMERABLE	Stool						MANDATORY	Multi Select
6	Result Time (min)	--	MEASURABLE	5-10	10-15	15-20	20-25	25-30	Minutes	MANDATORY, FILTER	
7	Ability to Evaluate Negative or Positive test result	--	ENUMERABLE	Yes					--	MANDATORY	
8	Sensitivity (%)	--	NUMERIC	90 -100					--	MANDATORY	
9	Specificity (%)	--	NUMERIC	90 -100					--	MANDATORY	
10	Contains an internal control line for the confirmation that the test has been performed correctly	--	BOOLEAN	Yes	No				--	GOLDEN	
11	Storage temperature	--	ENUMERABLE	2°C to 8°C	2°C to 30°C				--	MANDATORY, FILTER	
12	The supplier should ensure maintenance of recommended temperature during storage and transportation of Kit	--	ENUMERABLE	Yes					--	MANDATORY	
KIT CONTENTS											
1	Main items in test kit for performing the test	--	ENUMERABLE	Card					--	MANDATORY	
2	Clean, dry sterilized sample collection tube filled with sample diluent present with each kit	--	ENUMERABLE	Yes					--	MANDATORY	
3	Sample dropper provided	--	ENUMERABLE	Yes					--	MANDATORY	
4	Swab/spoon to collect specimen Provided	--	BOOLEAN	Yes	No				--	GOLDEN	
5	Dessicant to absorb moisture so that the Card/Strip do not get spoiled	--	BOOLEAN	Yes					--	MANDATORY	
6	Reactive and non reactive controls provided with each kit in adequate volume (minimum 10% of pack size)	--	ENUMERABLE	Yes	No				--	GOLDEN	
7	Packaging insert in English detailing the principle, components, methodologies, validity criteria, performance characteristics, bio-safety, limitations of assay, storage condition, manufacturing and expiry dates and methods of disposal Provided with each kit	--	ENUMERABLE	Yes					--	MANDATORY	
8	Other accessories and spares provided if any for standard pack in the kit	Must Declare	TEXT						--	MANDATORY	
PACKAGING											
1	Pack Size	--	ENUMERABLE	10 Tests Pack, 20 Tests pack, 25 Tests pack, 30 Tests pack, 50 Tests Pack, 100 Tests pack					--	GOLDEN	
2	The packing and labelling should be as per Drugs and Cosmetics Act, 1940	--	ENUMERABLE	Yes					--	MANDATORY	

3	Each card (cassette) should have space for patients particulars and date of the test	--	ENUMERABLE	Yes					--	MANDATORY	
4	The test kit should be packed in such a way that there is provision to conduct single test at a time	--	ENUMERABLE	Yes					--	MANDATORY	
5	Each test kit should be individually packed in a moisture proof pouches	--	ENUMERABLE	Yes					--	MANDATORY	
6	Test Kit container for housing all the items in the kit	--	ENUMERABLE	Yes					--	MANDATORY	

CERTIFICATIONS & REPORTS											
1	The kit should have approval of the statutory authority in its country of origin	--	ENUMERABLE	Yes					--	MANDATORY	
2	The Kit should be registered and licensed in India by DCGI (Proof of the same to be submitted to buyer on demand)	--	ENUMERABLE	Yes					--	MANDATORY	
3	Availability of valid drug license from competent authority defined under Drugs and Cosmetics Act, 1940 (Proof of the same to be submitted to buyer on demand)	--	ENUMERABLE	For Manufacturing	For Selling				--	MANDATORY, FILTER	Multi Select
4	Drug License Number	Must Declare	TEXT						--	MANDATORY	
5	Drug License Date	Must Declare	TEXT						--	MANDATORY	
6	Manufacturer certifications (Proof of the same to be submitted to the buyer on demand)	--	ENUMERABLE	GMP	WHO GMP				--	MANDATORY, FILTER	Multi Select
7	GMP/WHO GMP Certification Number	Must Declare	TEXT						--	MANDATORY	
8	GMP/WHO GMP Certification Date	Must Declare	TEXT						--	MANDATORY	
9	ISO 13485 Certified Manufacturer (Proof of the same to be submitted to buyer on demand in Certification is available)	--	BOOLEAN	Yes	No				--	MANDATORY, FILTER	
10	Product Certifications (Proof of the same to be submitted to buyer on demand)	--	ENUMERABLE	EU-CE	US-FDA				--	MANDATORY, FILTER	Multi Select
11	Availability of Test report from Central GOVT/NABL/ILAC accredited Lab to prove the conformity to the declared specification(proof of the same to be submitted to the buyer on demand)	--	ENUMERABLE	Yes					--	MANDATORY	
12	Availability of Performance Evaluation Report (Proof of the same to be submitted to the buyer on demand)	--	ENUMERABLE	Yes					--	MANDATORY	
13	Performance Evaluation Report issuing body	--	ENUMERABLE	National Insitute of Biological Sciences	Any other govt approved lab				--	MANDATORY, FILTER	
14	Name of the Performance Evaluation Report issuing body if other than National Institute of Biological Sciences	Must Declare	TEXT						--	MANDATORY	
SHELF LIFE											
1	Shelf Life (in months)	--	NUMERIC	24	30	36			Months	GOLDEN	
2	The product should not have passed more than 1/6 of the total shelf life at the time of dispatch to the consignee	--	Yes						--	MANDATORY	

(A. KESAVA RAO)
AD, GEM-SPV

(MANJU SHARMA)
DY.CEO

ACEO