

File No.GEM/TP/Shirodhara Yanthra System- Panchkarma/2019 (Version 3.0)							Draft T/P	GEM Version 3.0
Level 1 Category : Medical Equipment and Accessories and Supplies - UNSPS Code 42000000								
Level 2 Category : Ayurvedic Equipments and other Related Products - UNSPS Code 42990000								
Level 3 Category : Panchakarma Products and Accessories - UNSPS Code 42999900								
Level 4 Category : Shirodhara Yanthra with Pathra and Oil Heating System For Panchakarma - UNSPS Code 42999902							GEM Version 3.0	
S.NO	PARAMETER	HINT	TYPE OF FIELD	VALUES	UNIT	CATEGORY OF FIELD	REMARKS	
<b>GENERAL FEATURES</b>								
1	Product Description	--	ENUMERABLE	Shirodhara Yanthra System	--	MANDATORY		
2	Components	--	ENUMERABLE	Shirodhara Yanthra unit consist of Stand, Patra and Oil heating system	--	MANDATORY		
3	Utility	--	ENUMERABLE	For panchakarma of all age group	--	MANDATORY		
<b>SHIRODHARA STAND SPECIFICATIONS</b>								
1	Material of Shirodhara Stand	--	ENUMERABLE	Stainless steel	--	MANDATORY		
2	The stand should be made in fully adjustable design so that the Shirodhara vessel can be moved vertically and horizontally to any desired position using nylon bushed sliding brackets	--	ENUMERABLE	Yes	--	MANDATORY		
3	Ballasted for maximum stability during shirodhara	--	ENUMERABLE	Yes	--	MANDATORY		
4	Additional hook for hanging spare shirodhara vessel provided	--	ENUMERABLE	Yes	--	MANDATORY		
5	Shall be made using 38 mm dia pipes and 5 mm thick nylon bushes	--	ENUMERABLE	Yes	--	MANDATORY		
6	Provided with wheels for mobility, 3 nos star head bolts and 'U' type base	--	ENUMERABLE	Yes	--	MANDATORY		
7	Base Size (L x W) (±5%)	--	ENUMERABLE	60 cm x 40 cm	cm x cm	MANDATORY		
8	Stand Size (L x W) (±5%)	--	ENUMERABLE	180 cm x 60 cm	cm x cm	MANDATORY		
<b>SHIRODHARA PATHRA SPECIFICATIONS</b>								
1	Material	--	ENUMERABLE	Heat resistant corrosion free NPG Fiberglass	--	MANDATORY		
2	Provided with 0.25" brass fittings, stainless steel chains, fiberglass tube and CP ball valve control	--	ENUMERABLE	Yes	--	MANDATORY		
3	Should be suitable for use of any type of liquid including Thakradhara	--	ENUMERABLE	Yes	--	MANDATORY		
4	Made in shape of traditional pot in wood design	--	ENUMERABLE	Yes	--	MANDATORY		

5	Capacity	--	MEASURABLE	3.5			Litres	MANDATORY, FILTER	
6	Size (diameter x depth) (±5%)	--	ENUMERABLE	25 cm dia X 11 cm depth			--	MANDATORY	
<b>OIL HEATING SYSTEM SPECIFICATIONS</b>									
1	Should consist of traditional handmade 'White Bronze' vessel with an attached coiled heating element (for uniform oil heating ) and energy regulator to provide a steady flow of Dhara oil at constant heat	--	ENUMERABLE	Yes			--	MANDATORY	
2	Capacity of White Bronze vessel	--	MEASURABLE	3			Litres	MANDATORY	
3	Capacity of Energy regulator (watt)	--	ENUMERABLE	1200 W				MANDATORY	
4	Should be provided with CP ball valve, Brass rings and heavy duty Brass chain	--	ENUMERABLE	Yes			--	MANDATORY	
5	Control box should have MCB, heater indicator, SS Cowl and a bracket for fixing on to the Shirodhara stand	--	ENUMERABLE	Yes			--	MANDATORY	
6	Dimension of SS cowl (±5%)	--	ENUMERABLE	10 cm x 5 cm			--	MANDATORY	
7	Capacity	--	MEASURABLE	3.5			Litres	MANDATORY	
8	Size (dia x depth) (±5%)	--	ENUMERABLE	25 cm dia X 10 cm depth			--	MANDATORY	
9	A 'quick disconnect' connector is provided to remove the vessel	--	ENUMERABLE	Yes			--	MANDATORY	
<b>PACKING</b>									
1	The product should be packed with all its accessories in such a way that there is no transit damage takes place	--	ENUMERABLE	Yes			--	MANDATORY	
2	Packing Type (including user instructions)	--	ENUMERABLE	PACKING AS REQUIRED BY BUYER			--	MANDATORY	
<b>CERTIFICATIONS</b>									
1	ISO Certified Manufacturer (Proof of the same to be submitted to buyer on demand if ISO Certified)	--	ENUMERABLE	Yes	No		--	MANDATORY, FILTER	
2	Product Certifications (Proof of the same to be submitted to buyer on demand if certification is available)	--	ENUMERABLE	BIS	EU-CE	NA		MANDATORY, FILTER	

(A. KESAVA RAO)  
AD, GEM-SPV

(MANJU SHARMA)  
DY.CEO

ACEO