

Level 1 Category: Laboratory and Measuring and Observing and Testing Equipment - UNSPS Code 4100000

Level 2 Category: Laboratory and Scientific Equipment - UNSPS Code 4110000

Level 3 Category: Specimen Collection and Transport Containers and Supplies - UNSPS Code 41104100

Level 4 Category: Vacuum Blood Collection Tubes -UNSPS Code 41104107

Draft
T/P

GEM Version 3.0

S.NO	PARAMETER	HINT	TYPE OF FIELD	VALUES												UNIT	CATEGORY OF FIELD	REMARKS
GENERAL FEATURES																		
1	Product Description	--	ENUMERABLE	Vacuum Blood Collection Tubes												--	MANDATORY	
2	Application	--	ENUMERABLE	Biochemistry test, immunity test, and radioimmunoassay test												--	MANDATORY	
3	Sterilized	--	ENUMERABLE	Yes												--	MANDATORY	
4	Disposable	--	ENUMERABLE	Yes												--	MANDATORY	
PRODUCT TECHNICAL FEATURES																		
1	Material of Tube	--	ENUMERABLE	Glass	Platic, PET											--	GOLDEN	
2	Fine and transparent material	--	ENUMERABLE	Yes												--	MANDATORY	
3	Non Toxic, Leakproof and Crack Resistance	--	ENUMERABLE	Yes												--	MANDATORY	
4	Colored rubber stopper creating a vacuum seal inside the tube	--	ENUMERABLE	Yes												--	MANDATORY	
5	Negative air pressure present inside the tube	--	ENUMERABLE	Yes												--	MANDATORY	
6	Additive Present	--	ENUMERABLE	Yes	No											--	GOLDEN	
7	Type of Additive	--	ENUMERABLE	Caogulant	Anti coagulant	NA										--	MANDATORY, FILTER	
8	Additive Used	--	ENUMERABLE	K2 EDTA	K3 EDTA	Sodium Citrate	Sodium Heparin	Lithium Heparin	Sodium Fluoride	Sodium Oxalate	Potassium Oxalate	Clot Activator and Gel For Serum Separation	Clot Activator but no Serum Separation Gel	Acid Citrate Dextrose (ACD)	NA	--	GOLDEN	
9	Color of Cap	--	ENUMERABLE	Lavender or Purple	Light Green	Dark Green	Light Blue	Royal Blue	Gold	Yellow	Black	Gray	Pink	White	Red	--	MANDATORY, FILTER	
10	Color of Tube	--	ENUMERABLE	Transparent												--	MANDATORY	
11	Tube Size (mm x mm)	Must Declare	ENUMERABLE	8x120	13x75	13x100	16x100									mmxmm	MANDATORY	
12	Draw Volume of blood sample (ml)	--	MEASURABLE	1 to 5	5 to 7	8 to 10										ml	GOLDEN	
13	Shelf Life (in months)	Must Declare	NUMERIC													months	MANDATORY	
MARKING ON THE TUBE																		
1	Label on tube for Identification	--	ENUMERABLE	Yes												--	MANDATORY	
2	Label Type	--	ENUMERABLE	Paper	Paper - Crossmatch	See through										--	MANDATORY, FILTER	
PACKING																		
1	Packing	--	ENUMERABLE	Packed in a carton box												--	MANDATORY	
2	Number of tubes in a carton	--	NUMERIC	25	50	75	100									Nos	GOLDEN	
3	Price unit	--	ENUMERABLE	Each tube												--	MANDATORY	
CERTIFICATIONS & REPORTS																		
1	Submission of Test Report on form 39 or from Central Govt./NABL/ILAC accredited Lab to prove the conformity to declared specification at the time of supply	--	ENUMERABLE	Yes												--	MANDATORY	
2	Certification for Manufacturing Premises	--	ENUMERABLE	ISO	EU-CE	US-FDA	EU-CE&US-FDA	GMP								--	MANDATORY, FILTER	
3	Product conformity certificate is to be provided to the buyer at the time of supply	--	Yes													--	MANDATORY	

(A. KESAVA RAO)
AD, GEM-SPV(MANJU SHARMA)
DY.CEO

ACEO