

31	Pressure source	P	E	External pressurized air	External pressurized gas.									
32	Mobile trolley with IV pole should be provided	P	E	Yes										
33	Power supply	P	E	Single phase 230 V, 50 Hz										
34	UPS backup should be provided	P	E	Yes										
35	Type of microincision vitrectomy surgery(MIVS) system	P	E	Single entry system										
36	Availability to support microincision vitrectomy surgery(MIVS) options like 23 G & 25 G	P	E	Yes										
37	Availability of multifunctional foot switch with regular and dual linear control.	P	E	Yes										
38	System should be capable of Anterior vented gas forced infusion	P	E	Yes										
39	System should be capable to auotmated Silicon oil injection c	P	E	Yes										
40	Facility of auto fluid or air exchange	P	E	Yes										
41	Type of gas to auto fill	P	E	C3F8	SF6	Both C3F8 & SF6								
42	Facility for the extrusion of sub rental fluid provided	P	E	Yes										
43	Aavilability of MP3 Audio jack	P	E	Yes										
44	Facility of voice re-confirmation	P	E	Yes										
45	Availability of display to show parameters	P	E	Yes										
46	Availability to store various parameters	P	E	Yes										
47	Consumables Item													

65	The unit shall be capable of being stored continuously in ambient temperature of 0 - 50 degree centigrade and humidity of 15-90%	P	E	Yes										
66	The unit shall be capable of operating continuously in ambient temperature of 10 - 40deg C and relative humidity of 15-90%	P	E	Yes										
67	Certifications for product	C&R	E	US FDA	European CE	Both USFDA & European CE								
68	Certification no and date to be indicated Text	C&R	T											
69	Manufacturer should have ISO certification for quality standards	C&R	E	Yes										
70	ISO Certification No and date	C&R	T	Yes										
71	Copies of certifications to be furnished to buyer on demand after placement of order	C&R	E	Yes										
Mr.Ajmer Singh (TA)														
Shri A V Muralidhran (Dy CEO)														